


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PK-A-C-PN*

S.N.: *2632*

Date: *DEC. 2 85*



- a) Output signal amplitude:
0.70 ± 15 VOLTS
- b) Pulse width:
4.70 100 nsec
- c) Rise time:
(SEE FIGS 1 & 2)
≤ 100 psec
- d) Fall time:
≤ 2 nsec
- e) PRF:
MIN PW: 0.70 1 MHz
MAX PW: 0.70 20 kHz
- f) Jitter, stability: *(SEE FIGS 1 & 2)*
OK
- g) Prime power:
120/240 V
50-60 MHz

10 VOLTS / DIV
1.0 nsec / DIV
500 kHz

[Signature]