


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-M.-1402-PS-F*

S.N.: *2634*

Date: *OCT 18 85*



- a) Output signal amplitude:
0 to +20V
- b) Pulse width:
0 to 2 nsec
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 200 psec
- e) PRF:
0 to 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

16 V/DIV
1.0 nsec/DIV
1.0 MHz

120/240 V
50-60 Hz