


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-AM-14V2-PS-1*

S.N.: *2636*

Date: *OCT 18 '85*



- a) Output signal amplitude:  
*0 TO +20V*
- b) Pulse width:  
*0 TO 2 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 200 pSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*

*16 VOLTS/DIV.  
1.0 nSEC/DIV.  
1.0 MHz.*

*50-60 Hz.*  
*[Signature]*