


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AP-AN-1-C-PN*

S.N.: *2643*

Date: *OCT 21 '85*



- e) Output signal amplitude:  
*0 TO ±10.0.0V*
- f) Pulse width:  
*0 TO 4 μSEC*
- g) Rise time:  
*≤ 100 PSEC*
- h) Fall time:  
*≤ 135 PSEC*
- i) PRF:  
*0 TO 1 MHz*
- j) Jitter, stability:  
*OK*
- k) Prime power:

*10 VOLTS/DIV*  
*1.0 NS/DIV*  
*100 kHz*

*110/220V*  
*50-60 Hz*  
*[Signature]*