


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-B2-PS-05-N*

S.N.: *2652*

Date: *DEC 10 85*



- a) Output signal amplitude:
0 TO -50 VOLTS
- b) Pulse width:
20 TO 200 NSEC.
- c) Rise time:
≤ 3 NSEC
- d) Fall time:
≤ 3 NSEC.
- e) PRF:
0 TO 50 KHz.
- f) Jitter, stability:
OK.
- g) Prime power:
120/240 V
50-60 Hz.

10 VOLTS / DIV
50 NSEC / DIV
50 KHz.