


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-A-15-C-PN-05-M-D.

S.N.: 2656

Date: OCT 16 85



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
 ≤ 60 PSEC
- d) Fall time:
 ≤ 135 PSEC.
- e) PRF:
0 TO 1 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
1 NSEC/DIV
1 MHz

120/240V
50-60 Hz.