


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-P.*

S.N.: *2675*

Date: *OCT 30 85*



- a) Output signal amplitude:
0 TO +30 VOLTS
- b) Pulse width:
≤ 1 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

16 VOLTS/DIV
1.0 NSEC/DIV
500 KHz

+15 ± 1.0 VOLTS
150 mA (MAX)
[Signature]