


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-OT-P-PN*

S.N.: *2691*

Date: *NOV 4 85*



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0 TO 6 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
1.0 NSEC/DIV
5.0 MHz

120/240 V
50-60 Hz

h) OT (OFF-SET)
0 to ± 5 VOLTS
TO 50 Ω.