

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVC-TN-1

S.N.:

2698

Date:

OCT 11 85



- a) Output signal amplitude:
PULSE: -230 VOLTS
OFFSET: +30 VOLTS
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
L 2 NSEC
- d) Fall time:
L 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK.
- g) Prime power:

100 VOLTS PIV
10 NSEC PIV
1 KHz.

120/240 V
50-60 Hz

[Signature]