

PULSE GENERATOR
PERFORMANCE CHECK

Model:

AVO-7B-C-SNU

S.N.:

2708

Date:

NOV 19 85

a) Output signal amplitude:

0 TO +30 AMP

b) Pulse width:

0.5 TO 50 μ SEC

c) Rise time:

\leq 40 NSEC

d) Fall time:

\leq 40 NSEC

e) PRF:

MAX PW: 0 TO 50 HZ

f) Jitter, stability:

20 μ SEC OR LESS: 0 TO 300 HZ

OK

g) Prime power:

120/240 V

50-60 HZ

10 AMPS DIU
10 μ SEC DIU
 $R_L = 1.0 \Omega$
50 HZ

