


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-P*
S.N.: *2716*
Date: *OCT 22 85*



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
2 TO 100 NSEC.
- c) Rise time:
< 350 PSEC
- d) Fall time:
< 2 NSEC.
- e) PRF:
MIN PW: 0 TO 100 KHZ
MAX PW: 0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

2 NSEC/DIV
3.2 VOLTS/DIV
50 KHZ.

+15 ± 1.0 VOLT.
200 mA (MAX)

[Signature]