


PULSE GENERATOR
PERFORMANCE CHECK


Model: PI-V-HV3-C-PN-OS-M

S.N.: 2720

Date: NOV 8 85



(A) 100 VOLTS/DIV
10 NSEC/DIV
5 KHZ



(B) 100 VOLTS/DIV
1.0 NSEC/DIV
5 KHZ

- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 200 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 HZ