


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVH-5-1402-N*

S.N.: *2734*

Date: *NOV 29 85*



- a) Output signal amplitude:  
*0 TO - 30 VOLTS*
- b) Pulse width:  
*≤ 350 PSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*OK*

*32 VOLTS/DIV*  
*500 PSEC/DIV*  
*1.0 MHz*

*+15 ± 1.0 VOLT*  
*200 mA (MAX)*

*[Signature]*