


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AP-PU-2-W-GA-05-M*


S.N.: *2738*

Date: *OCT 15 85*



- a) Output signal amplitude:  
*0 TO +10 VOLTS*
- b) Pulse width:  
*20 TO 500 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 50 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 ± 1.0 VOLT*  
*200 mA (MPP)*

*A) 10 VOLTS/DIV  
10 NSEC/DIV  
50 KHz.*



*B) 10 VOLTS/DIV  
1.0 NSEC/DIV  
50 KHz.  
(RISE TIME):*