


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVN-1-C-PN-07-M.

S.N.: 2755

Date: NOV 28 85



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NSER
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 135 PSEC.
- e) PRF:
50 TO 250 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
1.0 NSER/DIV

120/240 V
50/60 Hz.

~~Signature~~
h) OFFSET 0 TO ± 5 VOLTS
TO 50 Ω .