


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2-PS-N-OS-M

S.N.: 2782

Date: DEC 13 85



- a) Output signal amplitude:  
0 TO -350 V
- b) Pulse width:  
0 TO 100 NSEC
- c) Rise time:  
≤ 1.5 NSEC
- d) Fall time:  
≤ 3 NSEC
- e) PRF:  
0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50 ~ 60 Hz

10 NSEC/DIV  
≈ 320 VOLTS/DIV  
1 KHz