


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-TN-1
S.N.: 2784
Date: NOV 25 85



- a) Output signal amplitude:
-230 V
- b) Pulse width:
5 TO 100 NSEC.
- c) Rise time:
≤ 2 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

100 VOLTS/DIV
10 NSEC/DIV
1 KHz

120/240 V
50-60 Hz

- h) OFFSET: 0 OR +30 V
- i) t_p : ≤ 100 NSEC.