


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVL-TN-1*  
S.N.: *2785*  
Date: *NOV 25 85*



- a) Output signal amplitude:  
*-230 V*
- b) Pulse width:  
*5 TO 100 NSEC*
- c) Rise time:  
*≤ 2 NSEC*
- d) Fall time:  
*≤ 3 NSEC*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*100 VOLTS/DIV      120/240 V*  
*10 NSEC/DIV      50-60 Hz*  
*1 KHz*

- h) OFFSET: *0 OR +30V*
- i) *t<sub>p</sub> ≤ 100 nsec.*

*[Signature]*