


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-AV-15-C-05-M-N-NP*

S.N.: *2795*

Date: *JAN 3 '86*



*5 VOLTS/DIV
1.0 NS/DIV
100 KHz*

- a) Output signal amplitude:
0 TO +5 V OUTS.
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 VOLTS,
50 - 60 Hz.*

[Signature]