


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-1-PS-P

S.N.: 2805

Date: JAN 28 86



- n) Output signal amplitude: +25 TO +100 AMP
- o) Pulse width: 25 NSEC
- p) Rise time: NA
- q) Fall time: NA
- r) PRF: 0 TO 1 KHz
- s) Jitter, stability: OK
- t) Prime power: 120 / 240 V

~ 60 AMPS/DIV
10 NSEC/DIV
1 KHz.

120 / 240 V
50-60 Hz
[Signature]