


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AN-2-C-N-PN-M*

S.N.: *2816*

Date: *NOV. 26 85*



-) Output signal amplitude:
0 TO ±10 VOLTS.
-) Pulse width:
5 TO 50 NSEC
-) Rise time:
≤ 100 PSEC
-) Fall time:
≤ 135 PSEC
-) PRF:
0 TO 50 KHz.
-) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
2.0 NSEC/DIV
50 KHz

120/240 V
50-60 Hz

[Signature]