


PULSE GENERATOR
PERFORMANCE CHECK

Model: *REP-N-15-C-N-PN-M*

S.N.: *2817*

Date: *NOV 26 85*



- d) Output signal amplitude:
0 TO ± 5 VOLTS.
- e) Pulse width:
0.2 TO 4.0 NSEC
- f) Rise time:
≤ 60 PSEC
- g) Fall time:
≤ 135 PSEC
- h) PRF:
0 TO 1 MHz.
- i) Jitter, stability:
OK
- j) Prime power:
*120/240 VOLTS,
50-60 HZ.*

*5 VOLTS/DIV
1.0 NSEC/DIV
1.0 MHz.*

[Signature]