


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM1-1-PS-P-PN-D*

S.N.: *2825*

Date: *JAN 21 86*



- d) Output signal amplitude:
0.70 ± 5 VOLTS
- e) Pulse width:
0.70 6 NSEC.
- f) Rise time:
≤ 100 PSEC
- g) Fall time:
≤ 135 PSEC.
- h) PRF:
0 TO 25. MHz
- i) Jitter, stability:
OK
- j) Prime power:

5 VOLTS/DIV
1.0 NSEC/DIV
5.0 MHz.

120/240 VOLTS
50-60 Hz

[Signature]