


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MM-1-PN*

S.N.: *2834*

Date: *JAN 24 '86*



- b) Output signal amplitude:  
*0 to ±5 VOLTS*
- c) Pulse width:  
*0 to 6 NS*
- d) Rise time:  
*≤ 100 PSEC*
- e) Fall time:  
*≤ 135 PSEC*
- f) PRF:  
*0 to 25 MHz*
- g) Jitter, stability:  
*OK*
- h) Prime power:

*5 VOLTS PIV  
100 NS  
5 MHz*

*+24 ± 1.0 VOLT  
450 mA (MAX)*