


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-17-MM1*  
S.N.: *2864*  
Date: *NOV 15 85*



- a) Output signal amplitude:  
*> +15 VOLTS.*
- b) Pulse width:  
*4 TO 100 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*MIN PW: 0 TO 1 MHz*  
*MAX PW: 0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*500 KHz*

*+15 ± 1.0 VOLT*  
*200 mA (MAX)*  
*[Signature]*