


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-15-RW1*

S.N.: *2865*

Date: *MAR 10 66*



a) Output signal amplitude:
> -0.2 VOLT

b) Pulse width:
≤ 0.2 NSEC

c) Rise time:
NA

d) Fall time:
NA

e) PRF:
0 TO 5 MHz

f) Jitter, stability:


OK

g) Prime power:

+15 ± 1.0 VOLT
50 mA (MAX)

h) input: *-1.0 VOLT,*
1.0 NSEC.

Ⓐ input: *0.5 VOLT/DIV*
0.5 NSEC/DIV



Ⓑ output: *0.5 VOLT/DIV* i) propagation delay:
0.5 NSEC/DIV *≤ 2 NSEC*