


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVMH-1-C-MIT-1*

S.N.: *2877*

Date: *JAN 10 86*



Ⓐ *1.0 NSER/DIV*  
*5 VOLTS/DIV*  
*13 MHz*



Ⓑ *100 PSEC/DIV*  
*5 VOLTS/DIV*  
*13 MHz*

- a) Output signal amplitude:  
*0.90 ± 10 VOLTS*
- b) Pulse width:  
*≤ 100 PSEC (FWHM)*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF: *0.90 15 MHz\**
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz.*

\* SPECIFICALLY  
DESIGNED TO  
OPERATE AT  
13 MHz.