


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PK-PP*

S.N.: *2907*

Date: *MAR 27 86*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
1 TO 50 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

1 NSEC / DIV
10 VOLTS / DIV
50 KHz

+15 ± 1.0 VOLTS
300 mA (MAX)