


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-1-PW-75-PN*


S.N.: *2913*

Date: *FEB 25 '86*



- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) Pulse width:
0.1 TO 5.0 USEC.
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC.
- e) PRF:
0 TO 1 KHZ.
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.

P_{out} *50 VOLTS/DIV*
1.0 USEC/DIV
R_L = 50 Ω



N_{out} *50 VOLTS/DIV*
1.0 USEC/DIV
R_L = 50 Ω

[Signature]