


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-102-P

S.N.: 2914

Date: FEB 25 1986



- a) Output signal amplitude:
 > +60 VOLTS
- b) Pulse width:
 2 NSCZ
- c) Rise time:
 NA
- d) Fall time:
 NA
- e) PRF:
 0 TO 1000 Hz
- f) Jitter, stability:
 OK
- g) Prime power:

≈ 32 VOLTS/DIV
2 NSCZ/DIV

+135V ± 5V
5mA (MAX)