


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-4V2-C-PW-M1*

S.N.: *2916*

Date: *max 31 86*



- a) Output signal amplitude:
0 TO ±20 VOLTS
- b) Pulse width:
5 TO 50 NS
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 150 PSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*2 NS/Div
10 VOLTS/Div
50 KHz*

*120/240V
50-60 Hz
h) offset: 0 TO
± 10 VOLTS TO
50 DIVS*