


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PR-3-PW-PS-PN-R*

S.N.: *2917*

Date: *MAR 10 1966*



a) Output signal amplitude:  
*0 TO ± 200 VOLTS*

b) Pulse width:  
*0.1 TO 100 USEC*


c) Rise time:  
*≤ 10 NSEC*

d) Fall time:  
*≤ 10 NSEC*

A)  $P_{out}$ : *50 VOLTS/DIV*  
*5 USEC/DIV*

e) PRF:  
*MAX PW: 0 TO 50 KHz*  
*MIN PW: 0 TO 1 KHz*

f) Jitter, stability:  
*OK*



g) Prime power:

*120/240 V*  
*50-60 HZ*

B)  $N_{out}$ : *50 VOLTS/DIV*  
*5 USEC/DIV*