


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-AD-C-M-N-PN*


S.N.: *2939*

Date: *FEB 28 '86*



10 NSER/DIV
10 VOLTS/DIV
50 KHZ

- a) Output signal amplitude:
0 TO ± 15 VOLTS
- b) Pulse width:
2 TO 5 NSER
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 2 NSER
- e) PRF:
0 TO 50 KHZ



1.0 NSER/DIV
10 VOLTS/DIV
50 KHZ

- f) Jitter, stability:
OK.
- g) Prime power:
120/240 V
50-60 HZ
[Signature]