


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVI-V-MV3-C-M-P-SNL1*

S.N.: *2952*

Date: *DEC 19 '85*



(A) *100 VOLTS/DIV*
10 NSEC/DIV
5 KHz



(B) *AS (A) BUT*
1.0 NSEC/DIV

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
0.90 100 NSEC
- c) Rise time:
≤ 200 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
- g) Prime power:
OK

120/240 V
50-60 Hz

[Signature]