


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-1-P
S.N.: 2957
Date: Feb 5 '86



- d) Output signal amplitude:
+25 TO +100 AMP
- e) Pulse width:
25 ± 5 NSEC
- f) Rise time:
N/A
- g) Fall time:
N/A
- h) PRF:
0 TO 1 KHz
- i) Jitter, stability:
OK
- j) Prime power:

≈ 60 AMP/DIV
10 NSEC/DIV
1 KHz.

+15 ± 1.0 VOLT.
200 mA (MAX)
[Signature]