


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-3-P*

S.N.: *2958*

Date: *MAY 30 '86*



- a) Output signal amplitude:  
*0 TO +15 VOLTS*
- b) Pulse width:  
*2 TO 15 USEC.*
- c) Rise time:  
*≤ 150 PSEC*
- d) Fall time:  
*≤ 600 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*2 USEC/DIV*  
*10 VOLTS/DIV*  
*5.0 MHz*

*+24 TO +28*  
*VOLTS,*  
*400 mA (MAX)*