


PULSE GENERATOR

PERFORMANCE CHECK


Model: *ANMB-2-PS-TRF-MA-PW*

S.N.: *2959*

Date: *MAY 30 86*



① *5 nsec/div*
10 volts/div
1.0 MHz



③ *AS ① BUT*
LOW TR, TF.

a) Output signal amplitude:

0 to ±20 VOLTS

b) Pulse width:

20 to 200 nsec

c) Rise time:

0.3 or 3 nsec

d) Fall time:

0.3 or 3 nsec

e) PRF:

0 to 5 MHz.

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz

[Signature]