


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-20C-PN-05*  
S.N.: *2967*  
Date: *DEC 18 85*



- d) Output signal amplitude:  
*0 TO ± 350 VOLTS*
- e) Pulse width:  
*5 TO 100 μSEC*
- f) Rise time:  
*≤ 1.5 NSEC.*
- g) Fall time:  
*≤ 3 NSEC*
- h) PRF:  
*0 TO 5 KHz.*
- i) Jitter, stability:  
*OK*
- j) Prime power:

*≈ 320 VOLTS/DIV  
10 NSEC/DIV  
1.0 KHz.*

*120/240V  
50-60 Hz*