


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PVI-U-1403-C-FCI-CA-N*

S.N.: *2974*

Date: *APRIL 3 66*



Ⓐ *50 VOLTS/DIV*
10 NSER/DIV
10 KHz



Ⓑ *AS Ⓐ BUT*
1.0 NSER/DIV

- a) Output signal amplitude:
0 TO -100 VOLTS
- b) Pulse width:
0 TO 100 NSER
- c) Rise time:
≤ 200 PSEZ
- d) Fall time:
≤ 3 NSER
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz