


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APV-AV-2-C-P*

S.N.: *2980*

Date: *MAR 13 '86*



- d) Output signal amplitude:  
*0 TO 410 VOLTS*
- e) Pulse width:  
*5 TO 50 NSEC*
- f) Rise time:  
*≤ 100 PSEC*
- g) Fall time:  
*≤ 135 PSEC*
- h) PRF:  
*0 TO 50 KHZ*
- i) Jitter, stability:  
*OK*
- j) Prime power:

*10 VOLTS/DIV*  
*5 NSEC/DIV*  
*50 KHZ*

*120/240V*  
*50 - 60 KHZ*  
