


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AN-15-C-P*

S.N.: *2981*

Date: *MAR 13 '86*



- a) Output signal amplitude:
0 TO +5 VOLTS.
- b) Pulse width:
0.2 TO 4 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*5 VOLTS/DIV
1.0 NSEC/DIV
100 KHz*

*120/240 V
50-60 Hz*