


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PNV-AV-15-N-C*

S.N.: *2985*

Date: *APRIL 10 86*



5 VOLTS/DIV
1.0 NS/DIV
1.0 MHz

- a) Output signal amplitude:
0 TO 5V
- b) Pulse width:
0.2 TO 4 NS
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

[Signature]