


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVMN-1-C-M-P*

S.N.: *2986*

Date: *APRIL 4 '86*



*5 VOLTS / DIV
1.0 NSEC / DIV
≈ 14 MHz*



*5 VOLTS / DIV
5.0 NSEC / DIV
≈ 50 MHz*

a) Output signal amplitude:

0 TO +5 VOLTS.

b) Pulse width:

0 TO 6 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 200 PSEC

e) PRF:

0 TO 50 MHz

f) Jitter, stability:

OK

g) Prime power:

120 / 240 V

50 - 60 Hz.