


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PUL-TN-1*

S.N.: *2988*

Date: *FEB 14 '86*



- a) Output signal amplitude:
a) - 230 VOLTS
- b) Pulse width:
*b) +30 V OFFSET.
5 TO 100 NSEC.*
- c) Rise time:
≤ 2 NSEC
- d) Fall time:
≤ 3 NSEC.
- e) PRF:
0 TO 5 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

*100 VOLTS/DIV.
10 NSEC/DIV.*

*120/240 V
50-60 Hz*

- h) PROPAGATION DELAYS:
≤ 100 NSEC.