


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-15-T*

S.N.: *3004*

Date: *APRIL 21 '86*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 4.0 μ SEC
- c) Rise time:
 \leq 60 PSEC
- d) Fall time:
 \leq 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 \pm 1.0 VOLTS.
300 mA (MAX)

500 PSEC/DIV
5.0 VOLTS/DIV
1.0 MHz

[Signature]