


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV4-S-1-P*  
S.N.: *3007*  
Date: *APRIL 21 '86*



- a) Output signal amplitude:  
*0 TO +10 VOLTS*
- b) Pulse width:  
*≤ 135 PSEZ*
- c) Rise time:  
*N/A*
- d) Fall time:  
*N/A*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*1.0 NSEZ/DIV*  
*1.0 VOLTS/DIV*  
*1.0 MHz*

*+15 ± 1.0 VOLT*  
*200 mA (MAX)*  
*[Signature]*