


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV0-2.W-C-P

S.N.: 3014

Date: MAY 16 86



- a) Output signal amplitude:  
0 TO + 10 AMPS
- b) Pulse width:  
2 TO 50 NSEC
- c) Rise time:  
≤ 1 NSEC
- d) Fall time:  
≤ 2 NSEC
- e) PRF:  
0 TO 20 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:

~ 7 AMPS/DIV  
5.0 NSEC/DIV  
10 KHZ.

120/240 V.  
50-60 HZ.