


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-3-C-N-PN-M*

S.N.: *3026*

Date: *APRIL 11 '86*



- a) Output signal amplitude:  
*0 TO ± 20 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 200 PSEC*
- d) Fall time:  
*≤ 300 PSEC*
- e) PRF:  
*0 TO 100 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240V*

*50-60 Hz*

*[Signature]*

*10 VOLTS/DIV*

*2.0 NS/DIV*

*1.0 MHz*