


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PMH-14V1-MMA1-P*

S.N.: *3027*

Date: *APRIL 15 86*



- a) Output signal amplitude:
> +100 VOLTS
- b) Pulse width:
≤ 1.0 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK.
- g) Prime power:

*100 VOLTS/DIV
1.0 NSEC/DIV
20 KHz*

*+15 ± 1.0 VOLTS
150 mA (MAX)*

[Signature]