


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-141-MA1*

S.N.: *3028*

Date: *APRIL 15 86*



- a) Output signal amplitude: *7.1 ± 1.0 VOLTS*
- b) Pulse width: *≤ 10 NSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 20 KHZ*
- f) Jitter, stability: *OK*
- g) Prime power:

100 VOLTS/DIV
1.0 NSEC/DIV
20 KHZ

+15 ± 1.0 VOLT
150 mA (MAX)
[Signature]